



Bid Number/बोली क्रमांक (बिड संख्या)<sup>:</sup> GEM/2024/B/5098326 Dated/दिनांक : 27-06-2024

#### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण		
Bid End Date/Time/बिड बंद होने की तारीख/समय	12-07-2024 12:00:00	
Bid Opening Date/Time/बिड खुलने की तारीख/समय	12-07-2024 12:30:00	
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)	
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Agriculture And Farmers Welfare	
Department Name/विभाग का नाम	Department Of Agricultural Research And Education (dare)	
Organisation Name/संगठन का नाम	Indian Council Of Agricultural Research (icar)	
Office Name/कार्यालय का नाम	Icar-nibsm, Raipur	
Total Quantity/कुल मात्रा	5	
ltem Category/मद केटेगरी	Tissue culture Racks for Plant Tissue culture in pre hardening Units (Q3)	
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No	
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No	
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC),OEM Authorization Certificate,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer	
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No	
Type of Bid/बिड का प्रकार	Two Packet Bid	
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days	
Inspection Required (By Empanelled Inspection Authority / Agencies pre- registered with GeM)	No	
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation	

EMD Detail/ईएमडी विवरण

Required/आवश्यकता		No
PBG Detail/ईपीबीजी विवरण		
Required/आवश्यकता		No
Splitting/विभाजन		
Bid splitting not applied/बोली विभाजन	लागू नहीं किया ग	गया.
111 Purchase Preference/एमआईआ	ई खरीद वरीयता	
MII Purchase Preference/एमआईआई	खरीद वरीयता	No
MSE Purchase Preference/एमएसई	खरीद वरीयता	
MSE Purchase Preference/एमएसई ख	रीद वरीयता	Yes
		<u> </u>
validated online through Udyam Reg Enterprises (MSEs) Order, 2012 data its subsequent Orders/Notifications Purchase preference, the bidder mu excluded from the purview of Public offering products manufactured by s for Services, the bidder must be the	istration porta d 23.03.2012 is ssued by conce st be the manu Procurement F come other OEI Service provid	ng valid Udyam Registration and whose credentials are I as defined in Public Procurement Policy for Micro and Small issued by Ministry of Micro, Small and Medium Enterprises and erned Ministry. If the bidder wants to avail themselves of the ufacturer / OEM of the offered product on GeM. Traders are Policy for Micro and Small Enterprises and hence resellers M are not eligible for any purchase preference. In respect of b ler of the offered Service. Relevant documentary evidence in respect of the offered product or service and Buyer will decid

validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1 + 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 <u>OM No.1 4 2021 PPD dated 18.05.2023</u> for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

## Tissue Culture Racks For Plant Tissue Culture In Pre Hardening Units ( 5 pieces )

Technical Specifications/तकनीकी विशिष्टियाँ

### \* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
General Parameter	Lighting Facility in Shelves	5

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य	
Electrical Parameter	Input Voltage	240.0, 220.0 (Volt)	
	Frequency	550.0 <b>(Hertz)</b>	
Constructional Parameter	Material	Mild Stee	

### Additional Specification Parameters - Tissue Culture Racks For Plant Tissue Culture In Pre Hardening Units ( 5 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)	
Lighting Lighting each Shelves should contain four fluorescent flicker free tube lights which are built in upper portion in shelves with separate control syst each Shelves. For four shelves- 16 and for 5 shelves- 20 tube lights be fitted. • Minimum one UV germicidal light fitted inside each rack. nos. of Spare tube lights of same fitting along with chowk/fully funct should be supplied.		
Easy Mobility	the whole rack is provided with four or more numbers of castor wheels at the bottom for easy mobility to any place with locking system. Minimum on year warranty for the time and tube lights fitting (Chowk/Frame etc.)	
Control panels	All control panels are fitted top of tissue culture rack/ with the individual selves. On/Off switch of fluorescent tube and UV tubes. All control panels are fitted separately in each rack with indicator light.	
timerIt must have 0-24 hrs photoperiodic timer to control the automatic line dark period.size and materialSize of each Shelves approx. 48'' x 18''to 24'' and distance between shelves: 16'' to 18''. Shelves should be made of SS/ MS with epoxy c material and each Shelves should have support for the tissue culture flask.		
		Bidders offering must also comply with the additional specification parameters mentioned above.

dders offering must also comply with the additional specification parameters mentioned above.

# Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Lalit Laxman Kharbikar	493225,ICAR-NATIONAL INSTITUTE OF BIOTIC STRESS MANAGEMENT, BARONDA	5	45

# Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्त

#### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

#### 2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document <u>Click here to view the file</u>.

# Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

- 1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
- 2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
- 3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
- 4. Creating BoQ bid for single item.
- 5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
- 6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
- 7. Floating / creation of work contracts as Custom Bids in Services.
- 8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for <u>attached categories</u>, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
- 9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
- 10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
- 11. Creating bid for items from irrelevant categories.
- 12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
- 13. Reference of conditions published on any external site or reference to external documents/clauses.
- 14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to

undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---